

WAC 448-13-020 Approval of breath test ((instruments))  
equipment. ((Pursuant to RCW 46.61.506, the DataMaster is the  
only breath test instrument approved by the state toxicologist  
as a device for the measurement of alcohol in a person's breath.  
A simulator filled with a certified simulator solution will be  
attached to each instrument to provide a known external standard  
as defined in WAC 448-13-030(13). The simulator used must be on  
the National Highway Traffic Safety Administration (NHTSA)  
conforming products list. Any agency, group, or individual  
seeking approval or certification from the state toxicologist  
for the use of other breath test instruments for evidential  
breath testing programs in the state of Washington should  
contact the state toxicologist at the address given in WAC 448-  
13-210.)) (1) Pursuant to RCW 46.61.506, the following  
instruments are approved for the quantitative measurement of  
alcohol in a person's breath:

(a) The DataMaster.

(2) Pursuant to RCW 46.61.506, the following thermometers  
are approved:

(a) Mercury in glass thermometers with a scale graduated in  
tenths of a degree measuring a range between 33.5 and 34.5  
degrees centigrade; and

(b) Digital thermometer system contained within the Guth  
2100 wet bath simulator.

[Statutory Authority: RCW 46.61.506. 95-20-025, § 448-13-020,  
filed 9/27/95, effective 10/28/95; 91-21-040, § 448-13-020,

filed 10/11/91, effective 11/11/91; 91-06-022, § 448-13-020,  
filed 2/26/91, effective 3/29/91.]